



**IAEA**

**2<sup>nd</sup> INTERLABORATORY COMPARISON FOR  
DEUTERIUM AND OXYGEN-18 ANALYSIS  
OF WATER SAMPLES**

**REPORT**

**Prepared by**

**J. Lippmann, M. Gröning and K. Rozanski**

**Isotope Hydrology Laboratory  
Agency's Laboratories Seibersdorf  
International Atomic Energy Agency**

**Vienna, September 1999**



## **Table of contents**

<b>1. Introduction.....</b>	<b>1</b>
<b>2. Presentation of the results.....</b>	<b>1</b>
<b>3. Statistical evaluation of the results.....</b>	<b>6</b>
<b>4. Summary of tables and figures.....</b>	<b>8</b>
<b>5. Conclusions.....</b>	<b>26</b>
<b>Acknowledgements .....</b>	<b>26</b>
<b>References .....</b>	<b>27</b>
<b>List of Tables .....</b>	<b>27</b>
<b>List of Figures .....</b>	<b>28</b>
<b>Annex I (List of participants) .....</b>	<b>29</b>
<b>Annex II (Reporting Sheet) .....</b>	<b>38</b>



## 1. INTRODUCTION

The IAEA Isotope Hydrology Laboratory organised in 1998/99 the 2nd interlaboratory comparison test for analytical laboratories engaged in routine analyses of hydrogen and oxygen stable isotope composition of water samples. This intercomparison exercise was carried out in the framework of the Analytical Quality Control Services (AQCS) programme of the International Atomic Energy Agency. The first interlaboratory comparison test of this kind carried out in 1995 was restricted to laboratories analysing, on a regular basis, precipitation samples collected in the framework of the Global Network "Isotopes in Precipitation" (GNIP), jointly operated by IAEA and WMO (Araguás-Araguás and Rozanski, 1995). The current test was open to all laboratories engaged in isotope analyses of water samples.

The interlaboratory comparison was announced in December 1998 on the Internet, via the ISOGEOCHEM news group. Moreover, invitation letters were sent out to all GNIP laboratories. Although the main objective of this exercise was to help the laboratories engaged in stable isotope analyses of water samples to assess their precision and accuracy for the range of  $\delta^{18}\text{O}$  and  $\delta^2\text{H}$  values normally observed in meteoric waters, it was decided that after completion of the exercise the analysed batches of water will be stored at the IAEA Isotope Hydrology Laboratory. They will serve as calibrated samples with the purpose to check the calibration of internal standards, and will be available on cost-free basis. It was therefore important to calibrate these materials with the smallest attainable uncertainty. To reach this goal, the IAEA specially invited seven laboratories known for the high quality of their analytical work to join the exercise and to provide the consensus values of the analysed waters. Four laboratories belonging to this group recently participated in the calibration of a successor material for VSMOW (M. Gröning – report in preparation).

More than one hundred laboratories from 43 countries indicated their willingness to participate in the exercise and received four water samples for analysis (IC-OH-1 to IC-OH-4, in the following indicated as OH-1 to OH-4). By the end of April 1999, 87 laboratories had returned the results to the IAEA (see list of participants in Annex 1.) Two laboratories - ID no. 112 and 134- reported only after the deadline: their results are listed, but have not been used for the evaluation. Each laboratory received a reporting sheet (see Annex II), requesting details of their sample preparation techniques and normalisation procedures. The laboratories were requested to report the  $\delta$ -values on the normalised VSMOW-SLAP scale, together with the overall uncertainty (one-sigma level).

The four water samples distributed among the participating laboratories cover the range of  $\delta^{18}\text{O}$  and  $\delta^2\text{H}$  values typical for the majority of natural waters. The samples were bottled from 60 l stainless steel storage barrels into 50 ml double capped brown glass bottles, serially numbered. Each laboratory received four samples with corresponding label numbers. These label numbers were further used throughout the exercise as laboratory identification (ID) numbers and are also used in the tables and graphs within this report. The ID numbers are not related to the list of participating laboratories given in Annex II.

## 2. PRESENTATION OF THE RESULTS

Tables 1 and 2 summarise the results of oxygen-18 and deuterium analyses of the intercomparison samples, as submitted by the laboratories participating in the exercise. The laboratories are identified by the assigned ID numbers (first column). The stated uncertainties are those quoted by the individual laboratories as one-sigma standard uncertainties. Although the laboratories have been asked to provide uncertainties assessed by a complete uncertainty budget or by the long-term standard deviation of internal laboratory standards (instructions given in the reporting sheet), most laboratories evaluated the uncertainty of the results only by using the standard deviation of individual measurements of the provided four samples. This has to be kept in mind for any quantitative assessment of the laboratory performance. All values are expressed in delta notation, normalised to the VSMOW-SLAP scale (Coplen et al., 1996).

Lab. ID no.	Preparation method	IC-OH-1		IC-OH-2		IC-OH-3		IC-OH-4	
		dO-18 ±1 sigma vs. VSMOW [‰]		dO-18 ±1 sigma vs. VSMOW [‰]		dO-18 ±1 sigma vs. VSMOW [‰]		dO-18 ±1 sigma vs. VSMOW [‰]	
10	CO2-eq	-0.04	0.05	-3.27	0.06	-8.60	0.04	-15.27	0.05
12	CO2-eq	0.20	0.40	-3.10	0.40	-8.60	0.40	-15.20	0.40
13	CO2-eq	-0.04	0.02	-3.28	0.02	-8.64	0.02	-15.27	0.02
14	CO2-eq	-0.20	0.50	-4.10	# 0.50	-9.00	0.50	-15.70	0.50
15	CO2-eq	-0.13	0.06	-3.33	0.05	-8.69	0.05	-15.30	0.05
16	CO2-eq	-0.22	x 0.06	-3.38	0.06	-8.74	0.06	-15.32	0.06
17	CO2-eq	-0.09	0.10	-3.37	0.10	-8.77	0.15	-15.49	0.20
19	CO2-eq	-0.02	0.10	-3.25	0.10	-8.67	0.10	-15.33	0.10
21	CO2-eq	-0.90	# 0.20	-3.40	0.20	-8.60	0.20	-15.00	0.20
22	CO2-eq	0.06	x 0.04	-3.27	0.05	-8.86	x 0.03	-15.76	x 0.04
24	CO2-eq	-0.10	0.06	-3.33	0.06	-8.60	0.06	-15.32	0.06
25	CO2-eq	-0.35	0.20	-3.52	0.20	-8.85	0.20	-15.35	0.20
26	CO2-eq	-0.05	0.07	-3.18	0.07	-8.69	0.07	-15.23	0.07
27	CO2-eq	-0.15	0.10	-3.30	0.10	-8.69	0.10	-15.31	0.10
28	CO2-eq	-0.17	0.10	-3.42	0.10	-8.80	0.10	-15.32	0.10
29	CO2-eq	0.00	0.30	-3.60	0.30	-8.60	0.30	-15.50	0.30
31	CO2-eq	-0.05	0.05	-3.25	0.05	-8.66	0.05	-15.34	0.05
32	CO2-eq	-0.08	0.05	-3.32	0.05	-8.74	0.05	-15.31	0.05
33	CO2-eq	-0.30	x 0.10	-3.50	x 0.10	-8.90	x 0.10	-15.60	x 0.10
34a	CO2-eq	-0.02	0.06	-3.11	x 0.07	-8.52	x 0.06	-15.20	0.05
35	CO2-eq	0.04	0.06	-3.17	x 0.05	-8.53	x 0.04	-15.11	0.12
36	CO2-eq	-0.09	0.04	-3.16	x 0.04	-8.61	0.05	-15.20	0.06
39	CO2-eq	-0.07	0.06	-3.39	0.07	-8.88	x 0.06	-15.67	x 0.10
42	CO2-eq	-0.24	0.10	-3.43	0.10	-8.86	0.15	-15.49	0.15
43	CO2-eq	-0.01	0.06	-3.16	x 0.04	-8.47	0.11	-15.26	0.07
44	CO2-eq	-0.07	0.09	-3.19	0.12	-8.58	0.14	-15.25	0.14
45	CO2-eq	0.51	# 0.33	-2.71	x 0.26	-7.80	# 0.15	-14.50	# 0.19
46	CO2-eq	-0.09	0.09	-3.30	0.09	-8.68	0.09	-15.36	0.09
47	CO2-eq	-0.20	0.10	-3.50	x 0.10	-8.80	0.10	-15.30	0.10
51	CO2-eq	0.00	0.22	-3.98	x 0.22	-8.92	0.22	-15.50	0.22
52	CO2-eq	-0.15	x 0.05	-3.30	0.05	-8.85	x 0.05	-15.45	x 0.05
53	CO2-eq	-0.08	0.04	-3.29	0.02	-8.67	x 0.01	-15.32	x 0.01
54	CO2-eq	-0.05	0.05	-3.22	0.05	-8.61	0.05	-15.29	0.05
55	CO2-eq	-0.01	0.06	-3.26	0.06	-8.61	0.06	-15.28	0.06
58	CO2-eq	-0.07	0.04	-3.27	0.04	-8.67	0.05	-15.30	0.06
61	CO2-eq	-0.14	0.08	-3.36	0.08	-8.67	0.08	-15.22	0.08
62	CO2-eq	-0.07	0.14	-3.36	0.08	-8.67	0.08	-15.34	0.08
64	CO2-eq	-0.22	0.10	-3.43	x 0.07	-8.77	0.07	-15.39	0.09
65	CO2-eq	-0.01	0.08	-3.20	0.08	-8.57	0.08	-15.16	0.08
67a	CO2-eq	-0.02	0.05	-3.25	0.05	-8.67	0.05	-15.32	0.03
68	CO2-eq	0.02	0.08	-3.21	0.08	-8.61	0.08	-15.27	0.08
69	CO2-eq	-0.26	x 0.07	-3.49	x 0.06	-8.95	x 0.06	-15.66	x 0.05

this table is continued on next page

Table 1, part 1/2: The results of  $\delta^{18}\text{O}$  analyses of four intercomparison samples (OH-1, OH-2, OH-3, OH-4) and the associated standard uncertainty, as reported by the laboratories participating in the interlaboratory comparison. Symbols (#) and (x) mark the results rejected in the first and second step of the statistical treatment, respectively; see also part 2 on next page.

Lab. ID no.	Preparation method	IC-OH-1 dO-18 ±1 sigma vs. VSMOW [‰]		IC-OH-2 dO-18 ±1 sigma vs. VSMOW [‰]		IC-OH-3 dO-18 ±1 sigma vs. VSMOW [‰]		IC-OH-4 dO-18 ± sigma vs. VSMOW [‰]	
continuation of the table from previous page									
71	CO2-eq	-0.13	0.14	-3.15	0.14	-8.49	0.13	-15.17	0.09
72	CO2-eq	0.01	0.15	-3.25	0.15	-8.53	0.15	-15.29	0.15
73	Hydrochloride	-0.20	0.20	-2.80	0.70	-8.00 #	0.50	-14.90	0.20
75	CO2-eq	-0.23 x	0.08	-3.34	0.08	-8.68	0.08	-15.36	0.08
78	CO2-eq	-0.08	0.05	-3.32	0.05	-8.72	0.05	-15.38	0.05
79	CO2-eq	-0.16 x	0.03	-3.43 x	0.04	-8.74	0.05	-15.44 x	0.05
82	CO2-eq	-0.08	0.10	-3.30	0.10	-8.68	0.10	-15.31	0.10
83	CO2-eq	-0.23 x	0.04	-3.47 x	0.08	-8.87 x	0.07	-15.43 x	0.05
85	CO2-eq	0.59 #	0.02	-2.65 x	0.04	-8.22 x	0.02	-15.05 x	0.03
86	CO2-eq	0.14 x	0.04	-3.11 x	0.04	-8.51 x	0.04	-15.21	0.04
88	CO2-eq	-0.18	0.12	-3.35	0.17	-8.61	0.17	-15.30	0.14
89	-	-	-	-	-	-	-	-	-
93	CO2-eq	-0.31 x	0.09	-3.47 x	0.08	-8.89 x	0.05	-15.54 x	0.03
94	-	-	-	-	-	-	-	-	-
96	CO2-eq	-0.11	0.10	-3.33	0.10	-8.71	0.10	-15.35	0.10
97	CO2-eq	0.04	0.05	-3.17 x	0.05	-8.57	0.07	-15.23	0.07
99	CO2-eq	-0.05	0.08	-3.29	0.08	-8.66	0.08	-15.27	0.08
101	CO2-eq	-0.06	0.11	-3.20	0.16	-8.31 x	0.11	-14.88 x	0.15
105	GC	-3.30 #	0.20	-6.60 #	0.20	-12.10 #	0.60	-18.00 #	0.10
106	CO2-eq	-0.81 #	0.25	-4.11 #	0.26	-9.32 #	0.15	-15.88 x	0.14
107	CO2-eq	-0.45	0.30	-3.72	0.30	-8.96	0.30	-15.54	0.30
108	CO2-eq	0.00	0.10	-3.27	0.10	-8.65	0.10	-15.32	0.10
109	CO2-eq	-0.14 x	0.02	-3.27	0.02	-8.77 x	0.04	-15.39 x	0.02
111	CO2-eq	-0.07	0.07	-3.31	0.07	-8.67	0.07	-15.32	0.07
113	CO2-eq	-0.16 x	0.05	-3.47 x	0.05	-8.66	0.11	-15.38	0.07
114	CO2-eq	-0.22	0.10	-3.45	0.10	-8.88 x	0.10	-15.48	0.10
115	CO2-eq	-0.01	0.04	-3.28	0.05	-8.63	0.02	-15.40 x	0.05
116	CO2-eq	-0.17 x	0.05	-3.36	0.07	-8.73	0.06	-15.35	0.06
117	CO2-eq	-0.17 x	0.05	-3.41 x	0.06	-8.73	0.07	-15.24	0.10
118	-	-	-	-	-	-	-	-	-
121	CO2-eq	-0.05	0.19	-3.25	0.19	-8.67	0.19	-15.33	0.19
122	CO2-eq	0.45 #	0.03	-2.87 x	0.03	-8.52 x	0.05	-15.37	0.08
123	CO2-eq	-0.17	0.15	-3.22	0.08	-8.51	0.08	-15.13	0.15
125	CO2-eq	-0.03	0.11	-3.23	0.11	-8.69	0.11	-15.51 x	0.11
126	CO2-eq	-0.10	0.10	-3.30	0.10	-8.60	0.10	-15.30	0.10
127	CO2-eq	0.07 x	0.05	-3.43 x	0.05	-8.78 x	0.05	-15.57 x	0.05
128	CO2-eq	-0.37 x	0.08	-3.58 x	0.08	-8.97 x	0.08	-15.58 x	0.08
129	CO2-eq	-0.08	0.05	-3.30	0.08	-8.65	0.09	-15.14	0.08
131	CO2-eq	-0.26 x	0.01	-3.45 x	0.02	-8.80 x	0.03	-15.47 x	0.06
132	CO2-eq	-0.13	0.05	-3.33	0.05	-8.72	0.05	-15.38	0.05
133	-	-	-	-	-	-	-	-	-
136	CO2-eq	-0.43 x	0.15	-3.64	0.20	-8.96	0.20	-15.54	0.21
137	CO2-eq	0.10 x	0.06	-3.20	0.04	-8.57	0.06	-15.32	0.07
These data arrived after the deadline and were not compiled									
112	CO2-eq	-0.17	0.14	-3.02	0.11	-8.70	0.14	-15.30	0.10
134	CO2-eq	-0.20	0.09	-3.1	0.05	-8.46	0.06	-15.36	0.08

Table 1, part 2/2: The results of  $\delta^{18}\text{O}$  analyses of four intercomparison samples (OH-1, OH-2, OH-3, OH-4) and the associated standard uncertainty, as reported by the laboratories participating in the interlaboratory comparison. Symbols (#) and (x) mark the results rejected in the first and second step of the statistical treatment, respectively.

Lab. ID no.	Preparation method	IC-OH-1		IC-OH-2		IC-OH-3		IC-OH-4	
		dD	$\pm 1$ sigma vs. VSMOW [‰]	dD	$\pm 1$ sigma vs. VSMOW [‰]	dD	$\pm 1$ sigma vs. VSMOW [‰]	dD	$\pm 1$ sigma vs. VSMOW [‰]
10	Zn, 500°C	-4.1	0.6	-31.0	0.7	-61.1	0.7	-109.0	0.9
12	Cr	-12.1 #	0.5	-37.9 #	0.5	-65.2 x	0.5	-111.0 x	0.5
13	Cr, 800°C	-2.7	0.7	-29.5	1.0	-60.1	0.8	-107.9 x	0.4
14	Cr	-3.6	0.2	-30.8	0.7	-61.0	1.0	-109.1	0.6
15	Zn, 700°C	-0.3 x	0.4	-30.7	0.9	-63.4 x	0.8	-112.9 x	1.5
16	Cr, 750°C	-4.1	0.8	-31.0	0.8	-61.3	0.8	-110.1	0.8
17	-	-	-	-	-	-	-	-	-
19	U	-4.1	0.5	-31.0	0.5	-61.1	0.5	-109.4	0.5
21	U, 850°C	-4.6 x	0.3	-31.9 x	0.3	-61.3	0.3	-109.7	0.3
22	-	-	-	-	-	-	-	-	-
24	-	-	-	-	-	-	-	-	-
25	Mn, 900°C	-3.4	1.5	-31.7	1.5	-63.5	1.5	-110.9	1.5
26	Zn	-3.1	0.9	-31.4	0.9	-61.5	0.9	-109.3	0.9
27	-	-	-	-	-	-	-	-	-
28	Zn	-4.0	1.0	-30.9	1.0	-61.3	1.0	-109.5	1.0
29	Zn, 500°C	-8.4	3.0	-33.3	3.0	-63.6	3.0	-108.3	3.0
31	Zn	-4.0	2.0	-33.0	2.0	-64.0	2.0	-112.0	2.0
32	-	-	-	-	-	-	-	-	-
33	Pt - eq.	-4.0	2.0	-32.0	3.0	-60.0	2.0	-110.0	2.0
34a	Zn	-4.3	2.7	-29.4	1.9	-62.7	2.0	-111.9	1.3
34b	Pt - eq.	-2.0	1.2	-28.9 x	0.8	-62.7	1.2	-110.9	1.5
35	Zn, 500°C	-18.0 #	2.8	-37.5 #	3.6	-66.5	4.4	-113.5	2.5
36	Pt - eq.	-3.9	0.4	-31.1	0.7	-61.9	0.7	-109.3	0.6
39	Pt - eq.	-2.0	1.9	-24.9 #	6.1	-58.3	6.8	-114.4	7.1
42	Zn	-3.5	1.0	-31.5	1.0	-62.0	1.0	-112.0	1.5
43	Zn	-2.2	0.9	-29.6	0.7	-59.9	1.5	-107.8	1.2
44	Cr, 910°C	-4.0	0.5	-30.7	0.5	-60.4	0.5	-108.2 x	0.5
45	Zn	-6.2	1.8	-32.0	3.3	-61.7	2.9	-108.6	2.1
46	-	-	-	-	-	-	-	-	-
47	Pt - eq.	-5.1	0.8	-32.2	0.8	-62.4	0.8	-110.9	0.8
51	-	-	-	-	-	-	-	-	-
52	Zn	-4.0	1.2	-32.0	1.2	-63.0	1.2	-108.0	1.2
53	Pt - eq.	-4.2	0.6	-30.9	0.6	-61.1	0.5	-109.2	0.4
54	Pt - eq.	-4.3	0.5	-31.3	0.5	-61.2	0.5	-109.6	0.5
55	-	-	-	-	-	-	-	-	-
58	U	-4.3	0.6	-31.0	0.7	-61.4	0.7	-109.7	0.5
61	Zn	-0.2 x	1.5	-28.6	1.5	-62.0	1.5	-109.8	1.5
62	-	-	-	-	-	-	-	-	-
64	Pt - eq.	-5.7 x	0.8	-32.1	0.8	-63.0 x	0.8	-110.4	0.9
65	Zn	-3.1	1.2	-31.1	1.2	-61.5	1.2	-109.0	1.2
67a	Zn, 480°C	-7.4 x	1.0	-32.4	1.0	-63.0	1.0	-110.9	1.0
67b	Cr, 850°C	-7.5 x	0.5	-32.7 x	0.5	-63.2 x	0.5	-111.2 x	0.5
68	Cr, 800°C	-3.0	0.5	-30.6	0.5	-60.4	0.5	-109.6	0.6
69	-	-	-	-	-	-	-	-	-

Table 2, part 1/2: The results of  $\delta^2\text{H}$  analyses of four intercomparison samples (OH-1, OH-2, OH-3, OH-4) and the associated standard uncertainty, as reported by the laboratories participating in the interlaboratory comparison. Symbols (#) and (x) mark the results rejected in the first and second step of the statistical treatment, respectively.



Lab. ID no.	Preparation method	IC-OH-1		IC-OH-2		IC-OH-3		IC-OH-4	
		dD	±1 sigma	dD	±1 sigma	dD	±1 sigma	dD	±1 sigma
vs. VSMOW [‰]									
continuation of table from the previous page									
71	Cr	-5.1	x 0.4	-31.7	0.5	-62.2	x 0.3	-110.3	x 0.3
72	Zn, 480°C	-6.6	1.5	-33.8	1.5	-63.3	1.5	-110.8	1.5
73	U	-4.8	x 0.2	-30.1	0.9	-62.7	1.1	-110.3	1.5
75	Pt - eq.	-4.5	0.8	-32.0	x 0.5	-61.7	0.9	-109.8	0.9
78	Zn, 480°C	-2.8	1.0	-29.7	1.0	-61.0	1.0	-110.3	1.0
79	Zn	-4.3	0.6	-31.8	0.7	-61.5	0.7	-110.1	0.9
82	Zn, 700°C	-4.1	1.0	-31.6	1.0	-61.5	1.0	-110.0	1.0
83	Zn, 450°C	-4.6	1.1	-31.1	1.1	-61.0	0.8	-109.4	1.2
85	Zn, 450°C	-9.2	x 2.2	-36.0	x 1.5	-67.0	x 1.7	-116.6	# 4.7
86	Zn	-4.2	0.5	-31.4	0.5	-61.3	0.5	-109.4	0.5
88	Zn	-6.7	1.8	-33.5	2.1	-65.4	2.1	-112.0	2.0
89	Zn	-4.6	1.3	-30.9	1.3	-61.6	1.3	-110.1	1.3
93	Pt - eq.	-4.1	1.6	-30.8	0.8	-60.3	1.8	-108.3	1.1
94	Zn, 500°C	-3.8	0.8	-31.2	0.5	-61.4	0.7	-109.7	1.0
96	Zn, 500°C	-1.8	2.0	-29.5	2.0	-58.8	2.0	-108.9	2.0
97	Pt - eq.	-5.4	1.0	-32.2	1.0	-62.3	1.5	-110.8	1.8
99	Zn, 500°C	-6.1	1.4	-32.9	1.4	-66.5	x 1.4	-115.5	x 1.4
101	Zn	-1.3	1.7	-32.3	1.6	-58.1	x 1.4	-115.3	x 1.1
105	GC	-24.0	# 0.8	-51.5	# 1.6	-85.6	# 2.3	-133.3	# 3.1
106	Pt - eq.	-6.3	5.7	-30.3	6.5	-58.7	4.3	-107.4	4.2
107	U	-3.0	3.0	-34.0	3.0	-62.4	3.0	-108.5	3.0
108	Pt - eq.	-3.8	1.0	-32.0	1.0	-60.8	1.0	-110.1	1.0
109		-4.0	0.6	-32.8	x 0.6	-61.6	0.6	-106.6	x 0.6
111	Pt - eq.	-3.0	1.0	-30.5	1.0	-61.4	1.0	-108.9	1.0
113	Zn	-4.4	1.1	-30.4	0.5	-60.3	1.3	-105.6	x 1.7
114	Pt - eq.	-3.9	1.0	-30.7	1.0	-60.3	1.0	-108.7	1.0
115	Cr	-6.0	x 0.1	-33.8	x 0.2	-64.9	x 0.3	-113.4	x 0.5
116	U	-3.9	0.6	-31.1	0.7	-60.7	0.6	-109.3	0.7
117	Zn	-4.5	0.8	-31.7	1.0	-61.5	1.0	-110.1	1.2
118	Zn	-4.0	2.0	-31.9	2.0	-72.1	# 2.0	-113.2	2.0
121	Pt - eq.	-4.0	2.0	-31.3	2.0	-60.1	2.0	-109.1	2.0
122	Pt - eq.	-2.5	x 0.3	-30.1	0.5	-60.7	0.6	-109.5	0.6
123	Zn, 500°C	-0.2	x 0.4	-28.9	x 0.2	-59.8	1.0	-107.8	2.5
125	Zn	-2.8	1.4	-31.1	1.4	-62.3	1.4	-109.4	1.4
126	Cr	-4.8	1.0	-31.8	1.0	-61.2	1.0	-109.0	1.0
127	Cr, 800°C	-0.8	x 1.0	-30.7	1.0	-60.5	1.0	-108.4	1.0
128	-	-	-	-	-	-	-	-	-
129	Zn	-1.9	x 0.4	-32.2	x 0.6	-60.4	0.7	-109.4	0.7
131	Zn	-4.0	0.4	-30.6	1.0	-60.8	0.4	-108.9	0.4
132	Pt - eq.	-3.4	1.1	-30.6	1.1	-60.5	1.1	-109.7	1.1
133	Zn, 480°C	-8.5	x 1.1	-31.2	1.7	-60.1	1.6	-103.1	# 1.9
136	Zn, 450°C	-2.7	1.2	-30.0	1.1	-58.6	x 1.1	-110.3	1.3
137	Zn, 450°C	-5.4	1.3	-32.3	0.9	-62.5	x 0.4	-110.9	1.0

These data arrived after the deadline and were not included in the compilation									
112	-	-	-	-	-	-	-	-	-
134	Zn	-0.6	1.0	-28.7	0.7	-59.9	0.8	-106.2	1.0

Table 2, part 2/2: The results of  $\delta^2\text{H}$  analyses of four intercomparison samples (OH-1, OH-2, OH-3, OH-4) and the associated standard uncertainty, as reported by the laboratories participating in the interlaboratory comparison. Symbols (#) and (x) mark the results rejected in the first and second step of the statistical treatment, respectively.

For oxygen-18 analyses, the  $\text{CO}_2$  equilibration method was used by almost all participating

laboratories. The conversion of water to hydrogen gas for deuterium measurements was done using one of the following methods: (i) reduction of water using granulate zinc (38 labs), (ii) reduction of water over hot uranium (6 labs), (iii) reduction of water over hot chromium (11 labs), and (iv) equilibration with H<sub>2</sub> using Pt as catalyst (18 labs) and (v) manganese reduction (1 lab).

### 3. STATISTICAL EVALUATION OF THE RESULTS

The statistical analysis of the submitted results was undertaken with two major objectives in mind:

- obtaining best estimates (consensus values) for the <sup>18</sup>O and <sup>2</sup>H content in the distributed samples as well as the standard uncertainty of these derived consensus values.
- assessing the performance of individual laboratories (precision and accuracy) with respect to the adopted consensus  $\delta^{18}\text{O}$  and  $\delta^2\text{H}$  values for the analysed samples.

To reach this goal, a two-stage statistical treatment, adopted in previous interlaboratory exercises of similar a nature (Rozanski et al., 1992; Araguàs-Araguàs and Rozanski, 1995), was applied separately for the subset of results originating from the selected group of laboratories and for the entire population of the submitted results.

#### The two-stage statistical treatment:

In stage I of the statistical treatment, obvious outliers were discarded based on the frequency distribution of values: after determination of the upper (H<sub>U</sub>) and lower (H<sub>L</sub>) quartiles, as well as the interquartile range (H<sub>U</sub>-H<sub>L</sub>), the values exceeding H<sub>U</sub> + 3.0·(H<sub>U</sub>-H<sub>L</sub>) and falling below H<sub>L</sub> -3.0·(H<sub>U</sub>-H<sub>L</sub>) were discarded. The values discarded in stage I are marked with #-symbol in Tables 1 and 2. The provisional mean after stage I is further used in the stage II outlier rejection procedure.

In stage II of the evaluation process, the remaining results were assessed through examination of the difference between each reported isotope result  $x$  and the provisional mean  $m$  of stage I, divided by the standard uncertainty  $s$  quoted by the given laboratory. The results for which the ratio  $|(x-m)|/s$  was larger than 2 were discarded. This procedure identifies those results which are not homogenous with the rest of the analysed population.

In the final step, the weighted average (eq. 1) was calculated by weighing the remaining individual results by the reciprocal of the quoted variance and the standard uncertainty was calculated with eq. 2:

$$X_w = \frac{\sum_{i=1}^n \frac{x_i}{s_i^2}}{\sum_{i=1}^n \frac{1}{s_i^2}} \quad (1)$$

$$stdev = \sqrt{\frac{\sum_{i=1}^n (\bar{x} - x_i)^2}{(n-1)}} \quad (2)$$

The estimated standard uncertainty of the mean (ese) was calculated according to the following formula

$$ese(X_w) = \frac{s_w}{\sqrt{\left(\sum_{i=1}^n \frac{1}{s_i^2}\right)}} \quad (3) \quad \text{with} \quad s_w^2 = \frac{\sum_{i=1}^n \frac{(x_i - X_w)^2}{s_i^2}}{n-1} \quad (4)$$

where  $n$  stands for the final number of results accepted after the second step of the data evaluation procedure.

Application of the two-stage statistical treatment on different data sets:

This two stage evaluation process was first applied on the subset of results submitted by the group of selected laboratories (Lab ID. No. 10, 13, 19, 32, 58, 109 and 111; ID No. 10 identifies the IAEA Isotope Hydrology Laboratory): Here the provisional *median* of the results after stage I is used in stage II instead of the provisional *mean*, since the number of laboratories  $n$  is smaller than 10. In stage II the weighted mean values (Table 3) and their standard uncertainties were calculated for each sample according to equations (1) and (2).

These weighted means of the subgroup of selected laboratories represent the adopted consensus values of the four samples (Table 5). The estimated standard uncertainty of the mean (ese) was calculated for each sample according to equation (3) and (4).

Results from selected laboratories in the intercomparison exercise

Sample	Original data median / range (reported values)	After removal of outliers (stage I & II) weighted mean $\pm$ stdev (accepted values)
delta O-18	[‰] vs. VSMOW	[‰] vs. VSMOW
OH - 1	-0.07 / 0.12 (7)	-0.05 $\pm$ 0.03 (6)
OH - 2	-3.27 / 0.07 (7)	-3.28 $\pm$ 0.03 (7)
OH - 3	-8.67 / 0.17 (7)	-8.65 $\pm$ 0.06 (6)
OH - 4	-15.31 / 0.12 (7)	-15.28 $\pm$ 0.03 (6)
delta D		
OH - 1	-4.1 / 1.6 (6)	-3.9 $\pm$ 0.9 (6)
OH - 2	-31.0 / 3.3 (6)	-30.8 $\pm$ 0.9 (5)
OH - 3	-61.2 / 1.5 (6)	-61.3 $\pm$ 0.3 (5)
OH - 4	-109.0 / 3.1 (6)	-109.4 $\pm$ 0.6 (4)

**Table 3:** The calculated standard deviations (stdev) are multiplied with an appropriate correction factor  $f$ , since the number of observations  $n$  is smaller than 10 (micro statistical group). The used correction factors  $f(n)$  are  $f(4) = 1.62$ ,  $f(5) = 1.42$ ,  $f(6) = 1.31$  and  $f(7) = 1.25$ .

Secondly, the evaluation process was applied to the entire population of the results (Table 4). The arithmetic mean value of the results remaining after stage I was adopted as a preliminary mean for stage II of the data evaluation process (m-value). Then, the weighted mean values and the associated standard uncertainties were calculated for each sample.

Results from all participating laboratories in the intercomparison exercise

Sample	Original data mean $\pm$ stdev (reported values)	After removal of outliers (stage I & II) weighted mean $\pm$ stdev (accepted values)
delta O-18	[‰] vs. VSMOW	[‰] vs. VSMOW
OH - 1	-0.10 $\pm$ 0.21 (80)	-0.11 $\pm$ 0.10 (61)
OH - 2	-3.33 $\pm$ 0.23 (80)	-3.30 $\pm$ 0.13 (59)
OH - 3	-8.68 $\pm$ 0.20 (80)	-8.69 $\pm$ 0.11 (55)
OH - 4	-15.34 $\pm$ 0.20 (80)	-15.33 $\pm$ 0.11 (58)
delta D		
OH - 1	-4.2 $\pm$ 2.0 (74)	-4.1 $\pm$ 1.3 (56)
OH - 2	-31.5 $\pm$ 1.8 (75)	-31.3 $\pm$ 1.0 (64)
OH - 3	-61.8 $\pm$ 2.1 (75)	-61.4 $\pm$ 1.4 (62)
OH - 4	-110.0 $\pm$ 2.1 (75)	-109.8 $\pm$ 1.4 (62)

**Table 4:** Summary statistics of the results obtained by all laboratories participating in the exercise. Weighted means were calculated according to eq.(1) and the standard deviation was calculated according to eq.(2).

Third, the evaluation process was performed again on the entire population, but this time the

adopted consensus value was used in the stage II outlier criterion instead of the provisional mean of all laboratories. This procedure allows to identify the outliers in the intercomparison exercise according to the adopted consensus values. The values discarded in this evaluation procedure are marked with x-symbol in Tables 1 and 2.

**Consensus value of the four intercomparison water samples**

Sample	consensus value	estimated standard uncertainty of the mean (ese) for the adopted consensus values
delta O-18	[‰] vs. VSMOW	[‰] vs. VSMOW
OH - 1	-0.05	0.01
OH - 2	-3.28	0.01
OH - 3	-8.65	0.02
OH - 4	-15.28	0.01
delta D		
OH - 1	-3.9	0.2
OH - 2	-30.8	0.2
OH - 3	-61.3	0.1
OH - 4	-109.4	0.2

**Table 5:** The adopted consensus  $\delta$ -values for the analysed water samples (column 2) derived from the results submitted by the group of seven selected laboratories. The estimated standard uncertainties of the mean (ese) for the adopted consensus values (column 3) were calculated according to eq.(3) and (4).

#### 4. SUMMARY TABLES AND FIGURES

Table 1 and 2 list the results of the  $\delta^{18}\text{O}$  and  $\delta^2\text{H}$  determinations performed on four intercomparison samples (OH-1, OH-2, OH-3, OH-4) and the associated standard uncertainty, as reported by the laboratories participating in the interlaboratory comparison. Symbols (#) and (x) mark the results rejected in the first and second step of statistical treatment, respectively, using the consensus values.

Table 3 summarises the results obtained for the selected group of seven laboratories, which were assigned to provide the consensus values for the analysed set of samples. In the second column, the median and the range of the results submitted by this group is reported. In the third column, the weighted mean (eq.1) and the associated standard uncertainty is reported, both calculated after removal of outliers (Stage I & II).

Table 4 provides analogous statistics for the entire population of the submitted results (including the group of seven selected laboratories).

In Table 5, the adopted consensus  $\delta$ -values for the analysed water samples are reported together with the estimated standard uncertainty of the mean (ese) for these consensus values, calculated using equations (1) and (2).

It is apparent from Tables 4 and 5 that the weighted mean values obtained from statistical evaluation of the entire population of the submitted results deviate systematically from the adopted consensus values. This shift towards more negative  $\delta$ -values amounts to ca.  $-0.04$  ‰ for  $\delta^{18}\text{O}$  and  $-0.3$  ‰ for  $\delta^2\text{H}$ . One possible explanation for this apparent discrepancy is discussed in the conclusions.

Lab. ID.Nr.	Preparation method	deviation with respect to the dO-18 consensus value [‰] vs. VSMOW				total deviation [‰]
		OH-1	OH-2	OH-3	OH-4	
10	CO2-eq	0.01	0.01	0.05	0.01	0.08
12	CO2-eq	0.25	0.18	0.05	0.08	0.56
13	CO2-eq	0.01	0.00	0.01	0.01	0.03
14	CO2-eq	-0.15	-0.82	-0.35	-0.42	1.74
15	CO2-eq	-0.08	-0.05	-0.04	-0.02	0.19
16	CO2-eq	-0.17	-0.10	-0.09	-0.04	0.40
17	CO2-eq	-0.04	-0.09	-0.12	-0.20	0.46
19	CO2-eq	0.03	0.03	-0.02	-0.05	0.12
21	CO2-eq	-0.85	-0.12	0.05	0.28	1.30
22	CO2-eq	0.11	0.01	-0.21	-0.48	0.81
24	CO2-eq	-0.05	-0.05	0.05	-0.04	0.19
25	CO2-eq	-0.30	-0.24	-0.20	-0.07	0.81
26	CO2-eq	0.00	0.10	-0.04	0.05	0.19
27	CO2-eq	-0.10	-0.02	-0.04	-0.03	0.18
28	CO2-eq	-0.12	-0.14	-0.15	-0.04	0.45
29	CO2-eq	0.05	-0.32	0.05	-0.22	0.64
31	CO2-eq	0.00	0.03	-0.01	-0.06	0.10
32	CO2-eq	-0.03	-0.04	-0.09	-0.03	0.19
33	CO2-eq	-0.25	-0.22	-0.25	-0.32	1.04
34a	CO2-eq	0.03	0.17	0.13	0.08	0.41
35	CO2-eq	0.09	0.11	0.12	0.17	0.49
36	CO2-eq	-0.04	0.12	0.04	0.08	0.28
39	CO2-eq	-0.02	-0.11	-0.23	-0.39	0.75
42	CO2-eq	-0.19	-0.15	-0.21	-0.21	0.76
43	CO2-eq	0.04	0.12	0.18	0.02	0.36
44	CO2-eq	-0.02	0.09	0.07	0.03	0.21
45	CO2-eq	0.56	0.57	0.85	0.78	2.76
46	CO2-eq	-0.04	-0.02	-0.03	-0.08	0.17
47	CO2-eq	-0.15	-0.22	-0.15	-0.02	0.54
51	CO2-eq	0.05	-0.70	-0.28	-0.21	1.24
52	CO2-eq	-0.10	-0.02	-0.20	-0.17	0.49
53	CO2-eq	-0.03	-0.01	-0.02	-0.04	0.10
54	CO2-eq	0.00	0.06	0.04	-0.01	0.10
55	CO2-eq	0.04	0.02	0.04	0.01	0.10
58	CO2-eq	-0.02	0.01	-0.02	-0.02	0.07
61	CO2-eq	-0.09	-0.08	-0.02	0.06	0.26
62	CO2-eq	-0.02	-0.08	-0.02	-0.06	0.18
64	CO2-eq	-0.17	-0.15	-0.12	-0.11	0.55
65	CO2-eq	0.04	0.08	0.08	0.12	0.32
67a	CO2-eq	0.03	0.03	-0.02	-0.04	0.12
68	CO2-eq	0.07	0.07	0.04	0.01	0.19
69	CO2-eq	-0.21	-0.21	-0.30	-0.37	1.10
this table is continued on next page						

Table 6, part 1/2: The deviations of the reported values with respect to the adopted consensus values for Oxygen-18. The last column contains the sum of the absolute values of the deviations (total).

Lab. ID.No	Preparation method	Deviation with respect to the 18 consensus value [‰] vs. VSMOW				dO. total deviation [‰]
		OH-1	OH-2	OH-3	OH-4	
continuation of table 6 from previous page						
71	CO2-eq	-0.08	0.13	0.16	0.11	0.48
72	CO2-eq	0.06	0.03	0.12	-0.01	0.21
73	Hydrochloride	-0.15	0.48	0.65	0.38	1.66
75	CO2-eq	-0.18	-0.06	-0.03	-0.07	0.34
78	CO2-eq	-0.03	-0.04	-0.07	-0.10	0.24
79	CO2-eq	-0.11	-0.15	-0.09	-0.16	0.51
82	CO2-eq	-0.03	-0.02	-0.03	-0.03	0.11
83	CO2-eq	-0.18	-0.19	-0.22	-0.15	0.74
85	CO2-eq	0.64	0.63	0.43	0.23	1.93
86	CO2-eq	0.19	0.17	0.14	0.07	0.57
88	CO2-eq	-0.13	-0.07	0.04	-0.02	0.26
93	CO2-eq	-0.26	-0.19	-0.24	-0.26	0.95
96	CO2-eq	-0.06	-0.05	-0.06	-0.07	0.24
97	CO2-eq	0.09	0.11	0.08	0.05	0.33
99	CO2-eq	0.00	-0.01	-0.01	0.01	0.04
101	CO2-eq	-0.01	0.08	0.34	0.40	0.83
105	GC	-3.25	-3.32	-3.45	-2.72	12.74
106	CO2-eq	-0.76	-0.83	-0.67	-0.60	2.86
107	CO2-eq	-0.40	-0.44	-0.31	-0.26	1.41
108	CO2-eq	0.05	0.01	0.00	-0.04	0.10
109	CO2-eq	-0.09	0.01	-0.12	-0.11	0.32
111	CO2-eq	-0.02	-0.03	-0.02	-0.04	0.11
113	CO2-eq	-0.11	-0.19	-0.01	-0.10	0.41
114	CO2-eq	-0.17	-0.17	-0.23	-0.20	0.77
115	CO2-eq	0.04	0.00	0.02	-0.11	0.17
116	CO2-eq	-0.12	-0.08	-0.08	-0.07	0.35
117	CO2-eq	-0.12	-0.13	-0.08	0.04	0.38
121	CO2-eq	0.00	0.03	-0.02	-0.05	0.10
122	CO2-eq	0.50	0.41	0.13	-0.09	1.12
123	CO2-eq	-0.12	0.06	0.14	0.15	0.47
125	CO2-eq	0.02	0.05	-0.04	-0.23	0.34
126	CO2-eq	-0.05	-0.02	0.05	-0.02	0.14
127	CO2-eq	0.12	-0.15	-0.13	-0.29	0.69
128	CO2-eq	-0.32	-0.30	-0.32	-0.30	1.24
129	CO2-eq	-0.03	-0.02	0.00	0.14	0.20
131	CO2-eq	-0.21	-0.17	-0.15	-0.19	0.72
132	CO2-eq	-0.08	-0.05	-0.07	-0.10	0.30
136	CO2-eq	-0.38	-0.36	-0.31	-0.26	1.31
137	CO2-eq	0.15	0.08	0.08	-0.04	0.34
These data arrived after the deadline and were not included in the compilation						
112	CO2-eq	-0.12	0.26	0.05	0.02	0.39
134	CO2-eq	0.25	0.18	0.19	0.12	0.74

Table 6, part 2/2: The deviations of the reported values with respect to the adopted consensus values for Oxygen-18. The last column contains the sum of the absolute values of the deviations (total).

Lab. ID.Nr.	Preparation method	Deviation with respect to the consensus value [‰] vs. VSMOW				total deviation [‰]
		OH-1	OH-2	OH-3	OH-4	
10	Zn, 500°C	-0.2	-0.2	0.2	0.4	1.0
12	Cr	-8.2	-7.1	-3.9	-1.6	20.8
13	Cr, 800°C	1.2	1.3	1.2	1.5	5.2
14	Cr	0.3	0.0	0.3	0.3	0.9
15	Zn, 700°C	3.5	0.1	-2.1	-3.5	9.2
16	Cr, 750°C	-0.2	-0.2	0.0	-0.7	1.1
19	U	-0.2	-0.2	0.2	0.0	0.6
21	U, 850°C	-0.7	-1.1	0.0	-0.3	2.1
25	Mn, 900°C	0.5	-0.9	-2.2	-1.5	5.0
26	Zn	0.8	-0.6	-0.2	0.1	1.7
28	Zn	-0.1	-0.1	0.0	-0.1	0.3
29	Zn, 500°C	-4.5	-2.5	-2.3	1.1	10.5
31	Zn	-0.1	-2.2	-2.7	-2.6	7.6
33	Pt	-0.1	-1.2	1.3	-0.6	3.2
34a	Zn	-0.4	1.4	-1.4	-2.5	5.7
34b	Pt - eq.	1.9	1.9	-1.4	-1.5	6.7
35	Zn, 500°C	21.9	-6.7	-5.2	-4.1	30.1
36	H2-eq.	0.0	-0.2	-0.6	0.1	1.0
39	H2 - eq.	1.9	5.9	3.0	-5.0	15.7
42	Zn	0.4	-0.7	-0.7	-2.6	4.4
43	Zn	1.7	1.2	1.4	1.6	5.9
44	Cr, 910°C	-0.1	0.1	0.8	1.2	2.2
45	Zn	-2.3	-1.2	-0.4	0.8	4.8
47	Pt - eq.	-1.2	-1.4	-1.1	-1.5	5.2
52	Zn	-0.1	-1.2	-1.7	1.4	4.5
53	Pt - eq.	-0.3	-0.1	0.2	0.3	0.9
54	Pt-eq.	-0.4	-0.5	0.1	-0.2	1.2
58	U	-0.4	-0.2	-0.1	-0.3	1.0
61	Zn	3.7	2.2	-0.7	-0.4	7.0
64	Pt-eq.	-1.8	-1.3	-1.7	-1.0	5.8
65	Zn	0.7	-0.3	-0.2	0.4	1.6
67a	Zn, 480°C	-3.6	-1.6	-1.7	-1.4	8.3
67b	Cr, 850°C	-3.6	-1.9	-1.9	-1.7	9.2
68	Cr, 800°C	0.9	0.2	0.9	-0.2	2.1

table 7 is contiued on next page

Table 7, part 1/2: Deviations of reported values with respect to the adopted consensus value for deuterium. The last column contains the sum of the absolute values of the departures (total).

Lab. ID.Nr.	Preparation method	Deviation with respect to the consensus value [‰] vs. VSMOW				total deviation [‰]
		OH-1	OH-2	OH-3	OH-4	
continuation of table 7 from previous page						
71	Cr	-1.2	-0.9	-0.9	-0.9	3.9
72	Zn, 480°C	-2.7	-3.0	-2.0	-1.4	9.1
73	U	-0.9	0.7	-1.4	-0.9	3.9
75	Pt-eq.	-0.6	-1.2	-0.4	-0.4	2.6
78	Zn, 480°C	1.1	1.1	0.3	-0.9	3.3
79	Zn	-0.4	-1.0	-0.2	-0.7	2.3
82	Zn, 700°C	-0.2	-0.8	-0.2	-0.5	1.7
83	Zn, 450°C	-0.7	-0.3	0.3	0.0	1.3
85	Zn, 450°C	-5.3	-5.2	-5.7	-7.2	23.4
86	Zn	-0.3	-0.6	0.0	0.0	1.0
88	Zn	-2.8	-2.7	-4.1	-2.6	12.2
89	Zn	-0.7	-0.1	-0.3	-0.6	1.7
93	Pt - eq.	-0.2	0.0	1.0	1.1	2.3
94	Zn, 500°C	0.1	-0.4	-0.1	-0.3	0.9
96	Zn, 500°C	2.0	1.3	2.5	0.6	6.4
97	Pt - eq.	-1.5	-1.4	-1.0	-1.4	5.4
99	Zn, 500°C	-2.2	-2.1	-5.2	-6.1	15.6
101	Zn	2.6	-1.5	3.2	-5.9	13.1
105	GC	-20.1	-20.7	-24.3	-23.9	89.0
106	Pt.-eq.	-2.4	0.5	2.6	2.0	7.5
107	U	0.9	-3.2	-1.1	0.9	6.1
108	Pt-eq.	0.1	-1.2	0.5	-0.7	2.4
109	Zn, 500°C	-0.1	-2.0	-0.3	2.8	5.3
111	Pt - eq.	0.9	0.3	-0.1	0.5	1.8
113	Zn	-0.5	0.4	1.0	3.9	5.7
114	Pt - eq.	0.0	0.1	0.9	0.7	1.8
115	Cr	-2.1	-3.0	-3.6	-4.0	12.7
116	U	-0.1	-0.2	0.6	0.1	1.0
117	Zn	-0.6	-0.9	-0.2	-0.7	2.4
118	Zn	-0.1	-1.1	-10.8	-3.8	15.8
121	Pt - eq.	-0.1	-0.5	1.1	0.3	2.1
122	Pt - eq.	1.4	0.8	0.6	-0.1	2.9
123	Zn, 500°C	3.7	1.9	1.5	1.6	8.7
125	Zn	1.1	-0.3	-1.0	0.0	2.4
126	Cr	-0.9	-1.0	0.1	0.4	2.4
127	Cr, 800°C	3.0	0.1	0.8	1.0	4.9
129	Zn	2.0	-1.4	0.9	0.0	4.3
131	Zn	-0.1	0.2	0.5	0.5	1.3
132	Pt - eq.	0.5	0.2	0.7	-0.3	1.7
133	Zn, 480°C	-4.6	-0.4	1.2	6.3	12.5
136	Zn, 450°C	1.2	0.9	2.7	-0.9	5.6
137	Zn, 450°C	-1.5	-1.5	-1.2	-1.5	5.7

This data arrived after the deadline and were not included in the compilation						
134	Zn	3.3	2.1	1.4	3.2	10.0

Table 7, part 2/2: Deviations of the reported values with respect to the adopted consensus value for deuterium. The last column contains the sum of the absolute values of the departures (total).



In Tables 6 and 7 the departures of reported values with respect to the adopted consensus values for oxygen-18 and deuterium are presented for all analysed samples. The last column in both tables contains the sum of the departures (absolute values) for the given isotope and for each participating laboratory.

Figures 1.1 to 1.8 show the  $\delta^{18}\text{O}$  and  $\delta\text{D}$  results with the associated standard uncertainties as provided by each laboratory for samples OH-1 to OH-4. The laboratory identification numbers were sorted according to the reported  $\delta$ -values: from more negative to more positive (S-shape plots). In order to keep all plots consistent with a constant range of  $\delta$ -values, some outliers were only partly plotted or not shown at all. The solid horizontal line indicates the adopted consensus value (Table 5, second column). The dashed line represents the mean value as derived for the entire population of the results after removal of outliers (Table 4, third column) with the standard deviation indicated by the two dotted lines.

Figures 2.1. to 2.4. show the  $\delta^{18}\text{O}$  -  $\delta^2\text{H}$  plots for the results submitted by the participating laboratories, for the samples OH-1 to OH-4, respectively. The indicated error bars represent the standard deviation of the initial data set of the two groups before the outlier rejection procedure: the two groups are 'all labs' (Table 4, second column: only lab 105 is already rejected as obvious outlier) and the 'seven selected labs' (Table 3, second column).

Figures 3.1 and 3.2. illustrate the distribution of the calculated departures of the measured  $\delta^{18}\text{O}$  and  $\delta^2\text{H}$  values with respect to the adopted consensus values. Laboratory identification numbers have been sorted according to increasing cumulative departures.

Figures 4.1 and 4.2 present the absolute cumulative departures for  $\delta^{18}\text{O}$  and  $\delta^2\text{H}$ , respectively, for each laboratory participating in the exercise.




Figure 1.1: S-shape plots of  $\delta^{18}\text{O}$  results for the sample OH-1 with values in ascending order versus the laboratory ID numbers. Vertical bars show the standard uncertainty ( $1\sigma$ ) quoted by each lab. The solid line represents the adopted consensus value. The dashed line indicates the weighted mean value as derived from the entire population of results after removal of outliers (Table 4, column3), dotted lines indicate its standard uncertainty ( $1\sigma$  level). Outliers are marked with open symbols.




Figure 1.2: S-shape plots of  $\delta^{18}\text{O}$  results for the sample OH-2 with values in ascending order versus the laboratory ID numbers. Vertical bars show the standard uncertainty ( $1\sigma$ ) quoted by each lab. The solid line represents the adopted consensus value. The dashed line indicates the weighted mean value as derived from the entire population of results after removal of outliers (Table 4, column3), dotted lines indicate its standard uncertainty ( $1\sigma$  level). Outliers are marked with open symbols.




Figure 1.3: S-shape plots of  $\delta^{18}\text{O}$  results for the sample OH-3 with values in ascending order versus the laboratory ID numbers. Vertical bars show the standard uncertainty ( $1\sigma$ ) quoted by each lab. The solid line represents the adopted consensus value. The dashed line indicates the weighted mean value as derived from the entire population of results after removal of outliers (Table 4, column3), dotted lines indicate its standard uncertainty ( $1\sigma$  level). Outliers are marked with open symbols.




Figure 1.4: S-shape plots of  $\delta^{18}\text{O}$  results for the sample OH-4 with values in ascending order versus the laboratory ID numbers. Vertical bars show the standard uncertainty ( $1\sigma$ ) quoted by each lab. The solid line represents the adopted consensus value. The dashed line indicates the weighted mean value as derived from the entire population of results after removal of outliers (Table 4, column3), dotted lines indicate its standard uncertainty ( $1\sigma$  level). Outliers are marked with open symbols.





Figure 1.5: S-shape plots of  $\delta^2H$  results for the sample OH-1 with values in ascending order versus the laboratory ID numbers. Vertical bars show the standard uncertainty ( $1\sigma$ ) quoted by each lab. The solid line represents the adopted consensus value. The dashed line indicates the weighted mean value as derived from the entire population of results after removal of outliers (Table 4, column3), dotted lines indicate its standard uncertainty ( $1\sigma$  level). Outliers are marked with open symbols.



61

Figure 1.6: S-shape plots of  $\delta^2\text{H}$  results for the sample OH-2 with values in ascending order versus the laboratory ID numbers. Vertical bars show the standard uncertainty (1  $\sigma$ ) quoted by each lab. The solid line represents the adopted consensus value. The dashed line indicates the weighted mean value as derived from the entire population of results after removal of outliers (Table 4, column3), dotted lines indicate its standard uncertainty (1  $\sigma$  level). Outliers are marked with open symbols.




Figure 1.7: S-shape plots of  $\delta^2H$  results for the sample OH-3 with values in ascending order versus the laboratory ID numbers. Vertical bars show the standard uncertainty ( $1\sigma$ ) quoted by each lab. The solid line represents the adopted consensus value. The dashed line indicates the weighted mean value as derived from the entire population of results after removal of outliers (Table 4, column3), dotted lines indicate its standard uncertainty ( $1\sigma$  level). Outliers are marked with open symbols.









Figure 1.8: S-shape plots of  $\delta^2\text{H}$  results for the sample OH-4 with values in ascending order versus the laboratory ID numbers. Vertical bars show the standard uncertainty ( $1\sigma$ ) quoted by each lab. The solid line represents the adopted consensus value. The dashed line indicates the weighted mean value as derived from the entire population of results after removal of outliers (Table 4, column3), dotted lines indicate its standard uncertainty ( $1\sigma$  level). Outliers are marked with open symbols.




**Figure 2.1:**  $\delta D$  versus  $\delta^{18}O$ : results of OH-1



**Figure 2.2:**  $\delta D$  versus  $\delta^{18}O$ : results of OH-2



**Figure 2.3:**  $\delta D$  versus  $\delta^{18}O$ : results of OH-3



**Figure 2.4:**  $\delta D$  versus  $\delta^{18}O$ : results of OH-4

# Deviation between reported and consensus dO-18 values of the four samples




Figure: 3.1

Laboratory identification number

## Deviation between reported and consensus dD values of the four samples




Figure: 3.2

Laboratory identification number





Figure 4.1 and Figure 4.2: Cumulative deviation from the consensus values for all four  $\delta^{18}\text{O}$  and  $\delta^2\text{H}$  results of each laboratory. Laboratories are sorted according to increasing cumulative deviation from the consensus values.



## 5. CONCLUSIONS

More than 80 laboratories performing routine  $\delta^{18}\text{O}$  and  $\delta^2\text{H}$  analyses of water samples participated in the 2<sup>nd</sup> interlaboratory comparison test organised by the IAEA. They represent a substantial portion of the entire analytical community engaged in this type of measurements. Therefore, it is believed that the results of this test properly reflect the current situation with respect to accuracy and interlaboratory precision of stable isotope analyses of water samples. The following conclusions can be drawn on the basis of the results presented in this report:

The apparent interlaboratory precision (one-sigma level), derived from the whole pool of the analysed results, is in the order of 0.11 ‰ for  $\delta^{18}\text{O}$  and 1.3 ‰ for  $\delta^2\text{H}$  (Table 4, third column). Thus, in first instance no significant improvement of this performance indicator could be detected with respect to the 1<sup>st</sup> interlaboratory comparison in 1995. But if one takes into account the different general conditions in 1995 and in 1999 - in 1995 all participants were asked to calibrate their system with the attached ampoules of VSMOW/GISP/SLAP, in 1999 this was omitted - therefore the results can be interpreted as a slight improvement of the overall performance: In 1999 about 50% of the participants calibrated their internal standards during the exercise.

The above quoted values are similar to typical standard uncertainties reported by the majority of laboratories engaged in routine analysis of water samples (0.1 ‰ for  $\delta^{18}\text{O}$  and 1.0 ‰  $\delta^2\text{H}$ ). Therefore, the conclusion drawn four years ago that these values set up practical limits for comparing the results originating from a large pool of different laboratories is still valid.

The apparent interlaboratory precision obtained for the group of seven selected laboratories was in the order of 0.04 ‰ for  $\delta^{18}\text{O}$  and 0.8 ‰ for  $\delta^2\text{H}$  (Table 3, third column).

Statistical analysis of the submitted results indicates that at least 25% of the laboratories participating in the exercise underestimate the standard uncertainty attached to their analyses and/or suffer from systematic effects of various nature (marked outliers).

No significant dependence of the obtained  $\delta$ -values on the type of the sample preparation procedure could be detected. Although all values obtained with the GC-method have been rejected, this might not be interpreted by statistical methods since only one laboratory reported to perform this preparation procedure.

The results for the entire population are not normal distributed, which indicates non random systematic effects in a substantial subgroup of laboratories.

A small but nevertheless significant discrepancy in the reported results between the group of selected laboratories and the entire population of participants is identified. This shift towards more negative  $\delta$ -values amounts to ca. -0.04 ‰ for  $\delta^{18}\text{O}$  and -0.3 ‰ for  $\delta^2\text{H}$  and pinpoints the importance of the applied calibration procedure (frequency, handling of primary standards, etc.) for the overall improvement of the precision and accuracy of the analysed water samples.

One possible explanation for the shift towards more negative values could be a slight evaporative enrichment of the used internal laboratory standards with time after the calibration against VSMOW (or evaporation of improperly stored VSMOW after initial opening of the glass ampoules, due to repeated use and/or leaky caps of a secondary storage container). It should be noted that evaporative loss of only 0.1% of the initial mass of the used internal standard or the VSMOW sample would be sufficient to explain the observed average isotope shift in the order of 0.04 ‰ for  $\delta^{18}\text{O}$  and 0.3 ‰ for  $\delta^2\text{H}$ .

## ACKNOWLEDGEMENTS

The efforts of all laboratories which participated in the 2<sup>nd</sup> IAEA interlaboratory comparison test and contributed with their results to the success of this exercise are highly appreciated. We would like to acknowledge the contribution of the group of selected laboratories which were instrumental in establishing the consensus  $\delta$  values for the analysed waters, and in particular the involvement of Liliana Andreescu, Mira Gattin, Manfred Jaklitsch and Herbert Tatzber as staff members of the IAEA Isotope Hydrology Laboratory in preparation, calibration and shipment of the distributed water samples.

## REFERENCES

Araguás-Araguás, L. and Rozanski, K. (1995). IAEA Interlaboratory Comparison for Deuterium and Oxygen-18 analysis of Precipitation Samples, IAEA Report, Vienna, 33 pp.

Coplen, T.B., P. De Bièvre, Krouse, H.R., Vocke, R.D., Gröning, M. and Rozanski, K. Ratios for light-element isotopes standardised for better interlaboratory comparison. *EOS*, July 2, 1996.

Rozanski, K., Stichler, W., Gonfiantini, R., Scott, E.M., Beukens, R.P., Kromer, B., Van der Plicht, J. (1992). The IAEA  $^{14}\text{C}$  intercomparison exercise, 34 (3), 506-519.

ISOGEOCHEM newsgroup: <http://geology.uvm.edu/geowww/isogeochem.html>

## LIST OF TABLES

Table 1: The results of  $\delta^{18}\text{O}$  analyses of four intercomparison samples (OH-1, OH-2, OH-3, OH-4) and the associated standard uncertainty, as reported by the laboratories participating in the interlaboratory comparison. Symbols (#) and (x) mark the results rejected in the first and second step of the statistical treatment, respectively.

Table 2: The results of  $\delta^2\text{H}$  analyses of four intercomparison samples (OH-1, OH-2, OH-3, OH-4) and the associated standard uncertainty, as reported by the laboratories participating in the interlaboratory comparison. Symbols (#) and (x) mark the results rejected in the first and second step of the statistical treatment, respectively.

Table 3: Summary statistics of the results obtained by the group of seven selected laboratories. The calculated standard uncertainties (stdev) are multiplied with an appropriate correction factor  $f$  to account for the low number of observations (microstatistics). The applied correction factors  $f(n)$  are  $f(4) = 1.62$ ,  $f(5) = 1.42$ ,  $f(6) = 1.31$  and  $f(7) = 1.25$ .

Table 4: Summary statistics of the results obtained by all laboratories participating in the exercise. Weighted means were calculated according to eq.(1) and the standard deviation was calculated according to eq.(2).

Table 5: The adopted consensus  $\delta$ -values for the analysed water samples (column 2) derived from the results submitted by the group of seven selected laboratories. The estimated standard uncertainties of the mean (ese) for the adopted consensus values (column 3) were calculated according to eq.(3) and (4).

Table 6: Differences between the  $\delta^{18}\text{O}$  values reported by the individual laboratories participating in the exercise and the consensus values adopted for the analysed water samples (OH-1 to OH-4). The last column contains the sum of all departures (absolute values) for each laboratory.

Table 7: Differences between the  $\delta^2\text{H}$  values reported by the individual laboratories participating in the exercise and the consensus values adopted for the analysed water samples (OH-1 to OH-4). The last column contains the sum of all departures (absolute values) for each laboratory.

## LIST OF FIGURES:

Fig.1.1.to 1.4.  $\delta^{18}\text{O}$  values for samples OH-1 to OH-4 in ascending order, as reported by the participating laboratories. Vertical bars represent the standard uncertainty (one-sigma) quoted by each laboratory. The solid line represents the adopted consensus value (Table 5, second column). The dotted line represents the weighted mean value as calculated from the whole set of laboratories with the standard uncertainty (one sigma level) indicated by the two dotted lines.

Fig.1.5 to 1.8.  $\delta^2\text{H}$  values for samples OH-1 to OH-4 in ascending order, as reported by the participating laboratories. Vertical bars represent the standard uncertainty (one-sigma) quoted by each laboratory. The solid line represents the adopted consensus value (Table 5, second column). The dotted line represents the weighted mean value as calculated from the whole set of laboratories with the standard uncertainty (one sigma level) indicated by the two dotted lines.

Figures 2.1. to 2.4. The  $\delta^{18}\text{O}$  -  $\delta^2\text{H}$  plots for the results submitted by the participating laboratories, for the samples OH-1 to OH-4, respectively. The indicated error bars represent the standard deviation (equ. 2) of the data of all labs (of the data of the selected labs, respectively) before the compilation procedure had been applied to the data sets.

Fig.3.1 and 3.2. Departures of the  $\delta^{18}\text{O}$  and  $\delta^2\text{H}$  values reported by the participating laboratories from the adopted consensus values. Laboratory identification numbers have been sorted according to increasing cumulative departures (absolute values).

Fig.4.1 and 4.2. Cumulative departures of the  $\delta^{18}\text{O}$  and  $\delta^2\text{H}$  values reported by the participating laboratories from the adopted consensus values (absolute values). Laboratories are sorted according to increasing value of this parameter.



## Annex I

### Participants in the 2<sup>nd</sup> IAEA intercomparison exercise of stable isotopes in precipitation, 1999

(listed in alphabetical order of their states)

Sue Wang  
Stable Isotope Laboratory  
School of Geosciences  
University of Wollongong  
Wollongong, NSW 2522  
Australia  
Tel.: +61-2-4221-4586  
Fax: +61-2-4221-4250  
Email: sue\_wang@uow.edu.au

Vitolds Gailitis  
CSIRO Land and Water  
Underwood Avenue  
Floreat Park  
Australia 6014  
Tel.: 641 8 93 33 62 79  
Fax: 61 8 9387 82111  
Email: vit@per.clw.csiro.au

Héctor O. Panarello  
Instituto de Geocronología y Geología Isotópica  
(INGEIS)  
Pabellon INGEIS, Ciudad Universitaria  
1428 - Ciudad de Buenos Aires  
Argentina  
Fax +54 1 783-3024  
Email: hector@ingeis.uba.ar,  
hector@mail.retina.ar

Albrecht Leis  
Joanneum Research Graz  
Institute of Hydrogeology and Geothermics  
Elisabethstr.16/II  
A-8010 Graz  
Austria  
Tel.: ++43 316 8761485  
Fax.: ++43 316 8761321  
Email: albrecht.leis@joanneum.ac.at

Wolfgang Papesch  
ÖFPZ Arsenal GmbH  
Faradaygasse 3  
1030 Wien  
Austria  
Tel: +43 1 79747 500  
Fax: +43 1 79747 587  
Email: papesch.w@arsenal.ac.at

Mira Gattin, Liliana Andreescu,  
Manfred Groening  
Isotope Hydrology Laboratory  
International Atomic Energy Agency  
Wagramer Strasse 5  
1400 Vienna  
Austria  
Tel.: +43 1 2600 21740  
Fax: +43 1 2600 7  
Email: M.Groening@iaea.org

Victoria Moreira, Marcelo Moreira  
Universidade de Sao Paulo  
Campus de Piracicaba  
Centro de Energia Nuclear na Agricultura CENA  
Avd. Centenario 303, Caixa Postal 96  
13400 Piracicaba  
Sao Paulo  
Brazil  
E-mail: reyna@pintado.ciagri.usp.br

Kurt Kyser  
Department of Geological Sciences  
Queen's University  
Kingston, ON K7L 3N6  
Canada  
Tel.: 613 545-6179  
Fax 613 545-6592  
Email: kyser@geol.queensu.ca

Richard Heemskerk, R.J. Drimmie  
Environmental Isotope Laboratory  
Department of Earth Sciences  
University of Waterloo  
Waterloo, Ontario  
Canada N2L 3G1  
Tel.: 519-888-4567 ext.-5838  
Fax: 519-746-0183  
Email: rkhmskrk@uwaterloo.ca

Stephen W. Taylor  
Stable Isotope Group  
University of Calgary  
2500, University Dr. N.W.  
Calgary Alberta  
T2N-1N4  
Canada  
Tel : 403 220 8268  
Fax: 403 220 7773  
Email : taylors@phas.ucalgary.ca

Leonard I. Wassenaar, Geoff Koehler  
Environment Canada  
Stable Isotope Hydrology and Ecology  
Laboratory  
11 Innovation Blvd.  
Saskatoon, SK  
Canada S7N 3H5  
Tel: (306) 975-5747  
Fax: (306) 975-5143  
Email: Len.Wassenaar@ec.gc.ca

Gilles St-Jean  
Isotope Laboratories  
University of Ottawa  
Dept of Earth Sciences  
140 Louis Pasteur, Ottawa, Ontario  
Canada, K1N 6N5  
Tel.: (613) 562-5800 ext. 6839  
Fax: (613) 562-5192  
Email: gstjean@science.uottawa.ca

Evelyn Aguirre D.  
Enviromental Isotope Seccion  
C.C.H.E.N.  
Amunategui 95  
Santiago  
Chile  
PO-Box 188-D  
Fax: 52-2-3646277  
Email: eaguirre@gopher.cchen.cl

Zonghu Zhang, Dong-Sheng Wang  
Institute of Hydrogeology and Engineering  
Geology  
Zhengding, Hebei  
P.R.China 050803  
Tel.: 0311-8029525 (Zhang, office)  
Email: globalg1@idt.net

Aly Islam Metwally Aly  
Central Lab. for Environmental Isotope  
Hydrology  
National Centre for Nuclear Safety and Radiation  
Control  
Egyptian Atomic Energy Authority  
3, Ahmed El-Zomor Street  
Nasr City, 11762  
P.O.Box: 7551, Cairo  
Egypt  
Fax: (202) 2740238  
Email: alyi45@yahoo.com

Antonio Matus, Julio Guidos  
División de Recursos Geotérmicos  
Comisión Ejecutiva Hidroeléctrica del Río  
Lempa  
Km 11 1/2 Carretera al puerto de la Libertad,  
Colonia Utila.  
Nueva San Salvador  
El Salvador, C.A.  
Tel.: 228-1400  
Fax: 228-0781  
E-mail: bmatus@hotmail.com

Tõnu Martma  
Laboratory of Isotope-Palaeoclimatology  
Institute of Geology at Tallinn  
Technical University  
Estonia Avenue 7  
Tallinn 10143  
Estonia  
Tel.: +372 645 4122  
Fax: +372 631 2074  
Email: martma@gi.ee

Eloni Sonninen  
Dating Laboratory  
University of Helsinki  
P.O.BOX 11  
00014 Helsinki  
Finland  
Tel.: +358-9-19123465  
Fax: +358-9-19123466  
Email: eloni.sonninen@helsinki.fi

Michel Stievenard  
CE Saclay  
LSCE/Bat 703  
91191 Gif sur Yvette cedex  
France  
Tel.: (33) 1 69 08 71 70  
Fax: (33) 1 69 08 77 16  
Email: misti@lsce.saclay.cea.fr

Agnès Noir, M. Dray  
Université Pierre et Marie Curie  
Centre de Recherches Géodynamiques  
47 Avenue de Corzent  
74203 Thonon Cedex  
France  
Email: dray@biogeodis.jussieu.fr,  
crg@biogeodis.jussieu.fr

Georges Lacrampe-Couloume  
Elf Exploration Production  
Isotope laboratory  
CSTJF L2/1047  
Avenue Larribau  
64018 PAU cedex  
France  
Tel.: + 33 5 59 83 44 13  
Fax ++ 33 5 59 83 45 66  
Georges.Lacrampe-Couloume@elf-p.fr

Christine Flehoc, Jean-Pierre Girard  
BRGM, SMN/ANA/ISO  
3 Avenue Claude Guillemin  
BP 6009  
F-45060 Orleanscedex 02  
France  
Tel.: 33 (0)2 38 64 34 13  
Fax : 33 (0)2 38 64 39 25  
Email : c.flehoc@brgm.fr , jp.girard@brgm.fr

Willibald Stichler  
Institut fuer Hydrologie  
GSF-Forschungszentrum  
Ingolstädter Landstrasse 1  
85764 Neuherberg  
Germany  
Email: stichler@gsf.de

Matthias Gehre, G.Strauch  
Centre for Environmental Research Leipzig-  
Halle  
Laboratory of Stable Isotopes  
Theodor-Lieser-Str. 4  
06120 Halle/Saale  
Germany  
Tel.: ++49 (345) 5585 240  
Fax.: ext. 559  
Email: gehre@ana.ufz.de

Christian Leibundgut, Paul Königer  
Institute of Hydrology  
University of Freiburg  
Fahnenbergplatz  
79098 Freiburg i.Br.  
Germany  
Tel: ++49 (0)761 203 3519  
Fax: ++49 (0)761 203 3594  
Email: koenigep@uni-freiburg.de

Klaus Simon  
Geochemisches Institut  
Universität Göttingen  
Goldschmidtstr. 1  
37077 Göttingen  
Germany  
Email: ksimon@gwdg.de

Axel Suckow, M.A. Geyh  
Geowissenschaftliche Gemeinschaftsaufgaben  
Joint Geoscientific Research Institute  
of the German State Geological Surveys  
Geochronology and Isotope Hydrology (S3)  
Stilleweg 2  
30655 Hannover  
Germany  
Tel.: +49 511 6432527  
Fax: +49 511 6433665  
Email: Axel.Suckow@bgr.de

Christel Facklam, Torsten Agemar  
Institute for Environmental Physics  
Im Neuenheimer Feld 229  
69120 Heidelberg  
Germany  
Tel.: (+49) 6221 54 6366  
Fax.: (+49) 6221 54 6405  
Email: fac@uphys1.uphys.uni-heidelberg.de,  
or C.Facklam@iup.uni-heidelberg.de

Manfred Schmitt  
Geochemische Analysen GCA  
Glückaufstrasse 50  
31319 Sehnde  
Germany  
Tel.: +49 5132 53579  
Fax: +49 5132 56346  
Email: GCA-Labor-MS@t-online.de

Torsten W. Vennemann  
Institute for Geochemistry  
University of Tuebingen  
Wilhelmstr. 56  
72076 Tuebingen  
Germany  
Tel: +49 (0)7071 297 4992  
Fax: +49 (0) 7071 29 5713  
Email: torven@uni-tuebingen.de

Willi A. Brand  
Isotope Laboratory  
Max-Planck-Institute for Biogeochemistry  
Tatzendpromenade 1a  
07745 JENA  
Germany  
Tel.: ++49 - 3641 - 643718  
Fax: ++49 - 3641 - 543710  
Email: wbrand@bgc-jena.mpg.de

Istvan Forizs, György Pantó  
Laboratory for Geochemical Research  
Hungarian Academy of Sciences  
H-1112 Budapest, Budaorsi ut 45  
Hungary  
Tel.: 36-1-319-3119  
Fax: 36-1-319-3145  
Email: forizs@sparc.core.hu

Sourendra K. Bhattacharya  
Physical Research Laboratory  
Ahmedabad 380 009  
India  
Tel: 91 79-462129  
Fax: 91 79-6560502  
Email: bhatta@prl.ernet.in

Suresh Navada  
Hydrology and Tracers Section  
Isotope Division  
Bhabha Atomic Research Centre  
Trombay, Mumbai 400085  
India  
Fax: 91-22-5505151  
Tel.: 91-22-5505050 Ext 2717.  
Email: snvnavada@apsara.barc.ernet.in

Zainal Abidin  
Hidrologi laboratorium  
Bidang Hidrologi dan industri  
Pusat Aplikasi Isotop dan Radiasi, Batan  
Jl. Cinere Ps Jumat  
PO Box 7002 JKSKL  
Jakarta 12070  
Indonesia  
Tel: 7659376  
Email: his45@bit.net.id

Ruth Yam, Aldo Shemesh  
Dept. of Environmental Sciences  
The Weizmann Institute of Sciences  
Rehovot, 76100  
Israel  
Tel: 972-8-9343429  
Fax: 972-8-9344124  
E-mail: cishemes@wiccmil.weizmann.ac.il

Mario E. Mussi  
Consiglio Nazionale delle Recherche  
International Institute for Geothermal Research  
Piazza Solferino 2  
56126 Pisa  
Italy  
Tel.: +39 (050) 41503 - 46069 - 41327  
Fax: +39 (050) 47055  
Email: gleone@dst.unipi.it

Naoki Kabeya, Nobuhito Ohte,  
Atsuko Sugimoto  
Forest Hydrology Laboratory  
Division of Environmental Science and  
Technology  
School of Agricultural Sciences  
Kyoto University  
Kyoto, 606-8502  
Japan  
Tel.: ++81 75 753-6093  
Fax: ++81 75 753-6088  
Email: kabeya@bluemoon.kais.kyoto-u.ac.jp

Atsuko Sugimoto  
Center for Ecological Research, Kyoto  
University  
Kamitanakami Hiranocho,  
Otsu 520-2113  
Japan  
Tel.: 077-549-8258 and/or -8200  
Fax 077-549-8201  
Email: atsukos@ecology.kyoto-u.ac.jp

Naohiro Yoshida  
Department of Environmental Science and  
Technology  
Interdisciplinary Graduate School of Science and  
Engineering  
Tokyo Institute of Technology  
4259 Nagatsuta, Midori-ku, Yokohama 226-8502  
Japan  
Tel: +81-45-924-5506 (office)  
Fax +81-45-924-5519  
Email: naoyoshi@depe.titech.ac.jp

Yong-Kwon KOH  
Geoenvironmental Sciences  
KAERI  
P.O. Box 105, Yusung, Taejon  
Korea 305-600  
Email: nykkoh@nanum.kaeri.re.kr

Andrew R. Campbell  
Professor of Geology  
Dept. of Earth and Environmental Science  
New Mexico Tech  
Socorro NM 87801  
U.S.A.  
Tel.: 505-835-5327  
Fax: 505-835-6436  
Email: campbell@mailhost.nmt.edu

Alvarado and Edith Cienfuegos,  
Pedro Morales  
Instituto de Geología, UNAM  
Circuito de la Investigación s/n  
Ciudad Universitaria.  
04510 México., D.F  
Mexico  
Tel. + (525) 622 43 25, Fax: -622 43 18  
Email: edithca@servidor.unam.mx

Mike K. Stewart  
GNS  
PO Box 31312  
Lower Hutt  
New Zealand  
Tel.: +644 570 4623  
Fax: +644 570 4657  
Email: m.stewart@gns.cri.nz

Ishaq Sajjad, M. Aznam Tasneem  
Radiation & Isotope Application Division  
PINSTECH  
P.O. Nilore  
Islamabad  
Pakistan  
Tel: 0092 51 9290261, Fax: 9290275  
Email: sajjad\_ishaq@yahoo.com

Rowena A. Isidro  
PNOC Energy Development Corporation  
EDRM-Central Chemistry Laboratory  
Gepscientific Department  
PNPC Complex, Merrit Road  
Fort Bonifacio, Makati City  
1201 Philippines  
Fax: +63-2-8446207 or +63-2-8152747  
Email: isidror@edc.energy.com.ph

Halina Mroz, Marek Dulinski,  
Kazimierz Rozansky  
University of Mining and Metallurgy  
Faculty of Physics and Nuclear Techniques  
Al. Mickiewicza 30  
30-059 Krakow  
Poland  
Tel.: 004812/6341996  
Fax: 004812/6340010  
Email: gorczyca@novell.ftj.agh.edu.pl,  
dulinski@novell.ftj.agh.edu.pl

Paula Maria Mimo Carreira Paquete  
Instituto Tecnológico e Nuclear  
Departamento de Química  
Estrada nacional N. 10  
2685 Sacavem  
Portugal  
Fax: + 351 1 9941455  
E-mail: carreira@itn1.itn.pt

Pokrovsky  
Geological Institute of Russian Academy of  
Science  
Pyzhevsky per., 7  
109017 Moscow  
Russia  
E-mail: pokrov@ginran.msk.su

Susan Waldron  
Life Sciences Community Stable Isotope Facility  
Scottish Universities Research and Reactor  
Centre  
East Kilbride G75 0QF  
Scotland  
Tel: (01355) 270135  
Fax: (01355) 229898  
Email: s.waldron@surrec.gla.ac.uk

Jan Kral, Ivan Rúčka, Juraj Michalko  
Department of Isotope Geology  
Geological Survey of Slovak Republic  
Mlynska dolina 1  
81704 Bratislava  
Slovak Republic  
Email: kral@gssr.sk, ruckai@eurotel.sk

Sonja Lojen  
Dept. of Environmental Science  
J. Stefan Institute  
Jamova 39  
1000 Ljubljana,  
Slovenia  
Tel.: +386 61 1885 393  
Fax: +386 61 1885 346  
E-mail: sonja.lojen@ijs.si

Siep Talma  
Quaternary Dating Research Unit  
Environmentek-CSIR  
PO Box 395  
0001 PRETORIA  
South Africa  
Tel.: ++ 27 12 841-3402  
Fax: ++ 27 12 349-1170  
Email: Stalma@csir.co.za

Verhagen, OHT Malinga  
Schonland Research Centre  
University of Witwatersrand  
P.O. Wits  
2050 Johannesburg  
South Africa  
Email: Verhagen@schonlan.src.wits.ac.za

Antonio Delgado Huertas  
Depto. Ciencias de la Tierra y Quimica  
Ambiental  
Estacion Experimental del Zaidin CSIC  
Prof. Albareda, 1  
18008 Granada  
Spain  
Tel: 34-958-121011  
Fax: 34-958-129600  
Email: antonio.delgado@eez.csic.es

Clemente Recio  
Stable Isotope Laboratory  
Fac. de Ciencias  
Univ. de Salamanca  
E 37008 - Salamanca  
Spain  
Tel.: (+34) 923 29 44 00 Ext. 1540  
Fax: (+34) 923 29 45 14  
Email: crecio@gugu.usal.es

Ramón Redondo-Ortega  
Laboratorio de Isótopos Estables. SIDI.  
Facultad de Ciencias, C-IX  
Universidad Autónoma de Madrid  
28049 Madrid  
Spain  
Tel.: 34-91- 397 48 98  
Fax: 34-91- 397 86 45  
Email: ramon.redondo@uam.es

Luis Araguas, Mari Fé Diaz Teijeiro, Antonio  
Plata  
Laboratorio de Hidrología Isotópica  
CETA/CEDEX Ministerio de Obras Públicas  
C/ Alfonso XII, 3  
28014 Madrid  
Spain  
Tel: +34 91 335 72 18  
Fax: +34 91 335 72 19  
E-mail: Luis.Araguas@cedex.es,  
Antonio.Plata@cedex.es

K.E.Anders Ohlsson  
Department of Forest Ecology  
Swedish University of Agricultural Sciences  
Petrus Laestadius väg  
S-901 83 Umeå  
Sweden  
Tel.: +46 (0)90-786 5892  
Fax +46 (0)90-786 7750  
Email: Anders.Ohlsson@sek.slu.se

Stefano Bernasconi  
Geologisches Institut  
ETH-Zentrum  
CH-8092 Zuerich  
Switzerland  
Tel. +41-1-632 3693  
Fax: +41-1-632 1080  
E-mail: stefano@erdw.ethz.ch

Markus Leuenberger  
Physikalisches Institut  
Abteilung Klima- und Umweltphysik  
Universitaet Bern  
Siderstrasse 5  
3012-Bern  
Switzerland  
Tel. +31 631 44 70  
Fax. +31 631 44 05  
Email: leuenberger@climate.unibe.ch

Rolf Siegwolf, Matthias Saurer  
Paul Scherrer Institute  
CH-5232 Villigen-PSI  
Switzerland  
Tel: +41 (0)56 310 27 86  
Fax: +41 (0)56 310 45 25  
Email: rolf.siegwolf@psi.ch,  
Matthias.Saurer@psi.ch

Jorge E. Spangenberg, J. Hunziker  
Stable Isotopes Laboratory  
Institute of Mineralogy and Petrography  
BFSH 2  
University of Lausanne  
CH 1015 Lausanne  
Switzerland  
Tel.: +41 21 692 4365 (office)  
Fax: +41 21 692 4305  
Email: Jorge.Spangenberg@imp.unil.ch

Chung-Ho Wang  
Stable Isotope Laboratory  
Institute of Earth Sciences, Academia Sinica  
#128, Sec.II, Yen-Chiu-Yuan Road  
Nankang, Taipei  
Taiwan 11529, R.O.C.  
Tel: 886-2-2783-9910  
Fax: 886-2-2651-1795  
Email: chwang@earth.sinica.edu.tw

Simon Kelly  
Food Authenticity Section  
CSL Food Science Laboratory  
Norwich Research Park  
Colney  
Norwich NR4 7UQ  
United Kingdom  
Tel.: 01603 590293  
Fax: 01603 501123  
Email: s.kelly@csf.gov.uk

Charles Belanger  
Analytical Laboratory  
PDZ Europa Ltd.  
Electra Way  
Crewe, Cheshire, CW1 6ZA  
United Kingdom  
Tel.: +44 1270-589398  
Fax +44 1270-589412  
E-mail: analytical@europa-uk.com

Paul Dennis  
Stable Isotope Laboratory,  
School of Environmental Sciences  
University of East Anglia  
Norwich NR4 7TJ  
United Kingdom  
Fax: 44 1 603-507719  
Email: p.dennis@uea.ac.uk

W G Darling  
British Geological Survey  
Maclean Building  
Wallingford  
Oxfordshire OX10 8BB  
United Kingdom  
Fax +1491 692345  
Email: wgd@wpo.nerc.ac.uk

Carol Arrowsmith  
NIGL  
British Geological Survey  
Nicker Hill  
Keyworth, Nottingham  
NG12 5GG  
United Kingdom  
Fax: +44 115 936 3302  
Email: c.arrowsmith@nigl.nerc.ac.uk

Michael Isaacs  
Isotope Unit, Dept. PRIS  
University of Reading  
Whiteknights  
Reading RG6 6AB  
United Kingdom  
Tel: +44(0)118-9875123x7974  
Fax: +44(0)118-9310279  
Email: slsisacs@rdg.ac.uk  
Andy Midwood  
Analytical Group  
Macaulay Land Use Research Institute  
Craigiebuckler  
Aberdeen, AB15 8QH  
United Kingdom  
Tel: +44 (0) 1224 318611  
Fax: +44 (0) 1224 311556  
Email: a.midwood@mluri.sari.ac.uk

Stephen Crowley  
Stable Isotope Laboratory  
Dept. of Earth Sciences  
University of Liverpool  
Brownlow Street  
PO Box 147  
Liverpool L69 3BX  
United Kingdom  
Email: sfcrow@liverpool.ac.uk

Lora Wingate  
University of Michigan  
Department of Geological Sciences  
Stable Isotope Laboratory  
425 East University  
1013 C.C. Little Building  
Ann Arbor, MI 48109-1063  
U.S.A.  
Tel.: (734) 763-0561  
Fax: (734) 763-4690  
Email: loraw@umich.edu

Robert Michener  
IRMS Laboratory Manager  
Boston University Stable Isotope Laboratory  
Department of Biology  
5 Cummington St.  
Boston, MA 02215  
U.S.A.  
Tel.: (voice) 617-353-6980  
Fax: 617-353-6340  
Email: michener@bio.bu.edu

Claudia I. Mora  
Department of Geological Sciences  
306 G & G Building

University of Tennessee  
Knoxville, Tennessee 37996-1410  
U.S.A.  
Fax: 423-974-2368  
Email: cmora@utk.edu

Tyler Coplen  
US Geological Survey  
Reston, VA 22092  
U.S.A.  
Email: tbcoplen@usgs.gov

David Dettman  
Isotope Hydrology Laboratory  
Department of Geosciences  
University of Arizona  
Gould-Simpson Bldg.  
Tucson, AZ 85721  
U.S.A.  
Tel.: (520) 621-4618  
Fax: (520) 621-2672  
Email: dettman@geo.arizona.edu

Bruce H. Vaughn  
Stable Isotope Laboratory  
Institute of Arctic and Alpine Research  
1560 30th Street  
Boulder, Colorado 80303  
U.S.A.  
Tel: (office): 303-492-7985  
(lab): 303-492-5495  
Fax: 303-492-6388  
Email: bruce.vaughn@colorado.edu

Douglas Introne  
Stable Isotope Laboratory  
Sawyer Environmental Research Center  
University of Maine  
Orono, ME 04469  
U.S.A.  
Tel.: (207) 581-2192  
Fax: (207) 581-3490  
Email: introne@maine.maine.edu

Andrea Lini  
Stable Isotope Laboratory  
Department of Geology  
University of Vermont  
Burlington, VT 05405  
U.S.A.  
Tel.: (802) 656 02 45  
Fax: (802) 656 00 45  
E-mail: alini@zoo.uvm.edu  
Steve Pelphrey  
Isotech Laboratories



1308 Parkland Court  
Champaign, IL 61821  
U.S.A.  
Fax: 217-398-3493  
Email: steve@isotech-labs.com

Mark Conrad  
Building 70A, Room 4418  
Center for Isotope Geochemistry  
E.O. Lawrence Berkeley National Laboratory  
Berkeley, CA 94720  
U.S.A.  
Tel: (510) 486-6141  
Fax: (510) 486-5496  
Email: MSConrad@lbl.gov

Matt Emmons  
Mountain Mass Spectrometry inc.  
634 Patty Drive  
Evergreen  
CO 80439  
U.S.A.  
Tel. and Fax: 303-674-5763  
Email: MattEmmons@aol.com

Peter Swart  
University of Miami  
RSMAS-MGG  
4600 Rickenbacker Causeway  
Miami, FL 33149  
U.S.A.  
Tel: 305-361-4103  
Fax: 305-361-4632  
Email: pswart@rsmas.miami.edu

William P. Patterson  
Department of Earth Sciences  
Syracuse University  
204 Heroy Geology Laboratory  
Syracuse NY 13244-1070  
U.S.A.  
Tel.: 315-443-3869 (office); 315-443-2937  
(isotope lab);  
Fax: 315-443-3363  
Email: wppatter@syr.edu

Dachun Zhang  
Isotope Laboratory  
Global Geochemistry Corp  
6919 Eton Ave  
Canoga Park, CA 91303  
U.S.A.  
Fax: 818-992-8940  
Email: globalg1@idt.net

Reed McEwan  
Stable Isotope Lab  
University of Minnesota  
301 Shepherd Labs  
100 Union St SE  
Minneapolis, MN 55455  
U.S.A.  
Tel.: 612.624.3348 and/or ext. 2338  
Fax: 612 625 3819  
Email: mcewa001@maroon.tc.umn.edu

## Annex II

### Reporting Sheet

sent to all participants of the 2<sup>nd</sup> IAEA intercomparison exercise  
of stable isotopes in precipitation, 1999  
as attachment to the shipment  
of the four water samples

**2<sup>nd</sup> IAEA INTERLABORATORY COMPARISON  
FOR STABLE ISOTOPE ANALYSES OF PRECIPITATION**

**Reporting Sheet**

Lab ID No.:  (Lab ID number is identical to the number stated on the sample bottles)

Participating  
Laboratory:

**SAMPLE PREPARATION DATA \***

Sample preparation technique (<sup>18</sup>O) + Mass Spectrometer used:

Sample preparation technique (<sup>2</sup>H) + Mass Spectrometer used:

\*) describe briefly the method employed for sample preparation (e.g. zinc reduction for <sup>2</sup>H and CO<sub>2</sub> equilibration for <sup>18</sup>O) and the mass spectrometer (model type, 2 or 3 Faraday-cup-system) .

**NORMALIZATION DATA \*\***

Normalization procedure for <sup>2</sup>H:

Date of last calibration using VSMOW/SLAP/GISP:

Normalization procedure for <sup>18</sup>O:

Date of last calibration using VSMOW/SLAP/GISP:

\*\*\*) describe briefly the procedure adopted in your laboratory to normalize the measured δ<sup>2</sup>H and δ<sup>18</sup>O values to the VSMOW-SLAP scale and state the isotopic composition of your calibrated laboratory standards.

**FINAL RESULTS**

SAMPLE	δ <sup>18</sup> O <sub>VSMOW/SLAP</sub> [‰]	Standard uncertainty <sup>#</sup> (one sigma)	Number of indiv. Analyses n(m) <sup>##</sup>	δ <sup>2</sup> H <sub>VSMOW/SLAP</sub> [‰]	Standard uncertainty <sup>#</sup> (one sigma)	Number of indiv. Analyses n(m) <sup>##</sup>	d-excess [‰]
IAEA-OH-1							
IAEA-OH-2							
IAEA-OH-3							
IAEA-OH-4							

**EVALUATION OF UNCERTAINTY COMPONENTS<sup>#</sup>**

- a)  by standard deviation of laboratory water standards.  
b)  by assessment of individual uncertainty components (state major ones):

Please note the remarks on Standard Uncertainty and Number of Analyses below before completing the form

#) Standard Uncertainty in table 'Final Results': This includes NOT ONLY the standard deviation of individual results of measurements of one sample. It rather means an assessment of all uncertainty components of the complete analysis procedure for the final result at your best knowledge (from the sample preparation: -e.g. temperature fluctuations of the equilibration bath- to the mass spectrometer: -variable isotope fractionation during gas transfer- and to the calculation: -assigned isotope values for the daily used laboratory water standard).

As a good approximation for the standard uncertainty in a laboratory one can use the standard deviation of individual results of a laboratory water standard over a reasonable period of time (two months, or more than 50 individual results). The standard uncertainty will often depend on the isotopic composition of the samples and may be larger for more negative  $\delta$ -values. Please be aware that reporting of the 'one sigma' standard uncertainty is requested, so - for statisticians and QA officers - the coverage factor  $k$  is set to 1.

##) Number of Analyses in table 'Final Results': Please state the number analyses  $n$  used for calculation of the mean and the number of analyses rejected internally  $m$  in brackets, (e.g. 9(1) means total of ten analyses performed, nine taken to calculate mean values, one result rejected internally).

### INDIVIDUAL ANALYSES

$\delta^{18}\text{O}_{\text{VSMOW/SLAP}}$ [‰], individual determinations				$\delta^2\text{H}_{\text{VSMOW/SLAP}}$ [‰], individual determinations				
	IAEA- OH-1	IAEA- OH-2	IAEA- OH-3	IAEA- OH-4	IAEA-OH- 1	IAEA- OH-2	IAEA- OH-3	IAEA- OH-4
1.								
2.								
3.								
4.								
5.								
6.								
7.								
8.								
9.								
10.								
add more								

Laboratory Manager.....

Date:.....

Please send this reporting sheet (electronic version strongly preferred) to:

**Manfred Gröning**  
**Head, Isotope Hydrology Laboratory**  
**International Atomic Energy Agency**  
**P.O.Box 100 A-1400 Vienna AUSTRIA**  
**Phone: +43-1-2600-21740/21766 Fax: +43-1-2600-7**  
**e-mail (SMTP) : rialihl@iaea.org**